
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007720
8:03 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:32
STERILIZING 4:36
136.0 C 220kPa 4:36
136.0 C 220kPa 5:06
136.4 C 219kPa 5:36
136.2 C 221kPa 6:06
136.3 C 223kPa 6:36
136.4 C 223kPa 7:06
136.3 C 220kPa 7:36
136.2 C 221kPa 8:06
136.2 C 221kPa 8:06
VENTING 8:06
AIR DRYING 9:00
CYCLE COMPLETE 9:10

Digital Signature #
4EED0EBFC42B2DB8
