
Midmark M11 - v1.0.5
Total Cycles: 3658

Name: RITTER2
Sterilizer ID: V2313228
05/30/2024 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.3 F 0.0
4:00 75.4 F 0.2
6:00 103.1 F 1.5
8:00 138.1 F 4.0
10:00 163.7 F 6.9
12:00 184.7 F 10.1
14:00 201.5 F 13.6
16:00 221.2 F 6.7
18:00 227.9 F 10.6
20:00 227.4 F 7.6
22:00 237.2 F 10.8
24:00 233.3 F 7.8
26:00 246.4 F 13.8
28:00 257.2 F 19.6
30:00 267.0 F 25.7
32:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 273.0 F 29.9
2:00 271.7 F 29.1
3:00 271.7 F 29.6
4:00 272.1 F 29.7

Min 271.1 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 221.4 F 0.0
10:00 185.9 F 0.0
15:00 180.0 F 0.0
20:00 175.1 F 0.0
25:00 172.1 F 0.0
30:00 170.8 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:35
CYCLE COMPLETE

05/30/2024 09:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____