

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1519

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2022 05:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 157.5 F 2.3  
4:00 179.2 F 6.3  
6:00 196.2 F 11.2  
8:00 214.3 F 4.1  
10:00 228.1 F 9.1  
12:00 237.2 F 10.0  
14:00 241.0 F 11.4  
16:00 238.2 F 9.7  
18:00 252.1 F 16.4  
20:00 263.7 F 23.2  
22:00 271.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.7 F 29.4  
2:00 271.7 F 28.6  
3:00 272.5 F 29.3  
4:00 272.4 F 29.1

Min 271.3 F 28.4  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 232.9 F 0.0  
10:00 242.7 F 0.0  
15:00 234.0 F 0.0  
20:00 227.2 F 0.0  
25:00 229.8 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

01/13/2022 07:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02