
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002565
10:47 07/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:04
STERILIZING 3:06
136.0 C 221kPa 3:06
136.4 C 224kPa 3:36
136.4 C 223kPa 4:06
136.4 C 223kPa 4:36
136.4 C 224kPa 5:06
136.3 C 223kPa 5:36
136.5 C 226kPa 6:06
136.2 C 224kPa 6:36
136.2 C 224kPa 6:36
VENTING 6:36
AIR DRYING 7:28
CYCLE COMPLETE 13:29

Digital Signature #
450EC2B464458321
