
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000655
10:50 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:01
STERILIZING 4:09
136.0 C 227kPa 4:09
136.1 C 231kPa 4:39
136.4 C 231kPa 5:09
136.3 C 232kPa 5:39
136.4 C 233kPa 6:09
136.3 C 230kPa 6:39
136.4 C 233kPa 7:09
136.3 C 232kPa 7:39
136.3 C 232kPa 7:39
VENTING 7:39
AIR DRYING 8:34
CYCLE COMPLETE 8:36

Digital Signature #
87EE730E7601F0ED
