
Midmark M11 - v1.0.5
Total Cycles: 2728

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 161.0 F 0.0
4:00 177.2 F 0.8
6:00 202.4 F 5.6
8:00 219.2 F 10.8
10:00 225.5 F 6.4
12:00 237.0 F 12.1
14:00 237.3 F 10.2
16:00 236.5 F 9.2
18:00 251.1 F 16.4
20:00 263.4 F 23.5
22:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 28.9
3:00 272.7 F 29.8
4:00 272.0 F 29.1

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 213.9 F 0.0
10:00 220.0 F 0.0
15:00 220.1 F 0.0
20:00 223.1 F 0.0
25:00 229.9 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

08/03/2023 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____