
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001828
14:47 11/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:40
STERILIZING 3:44
136.5 C 224kPa 3:44
136.4 C 225kPa 4:14
136.4 C 225kPa 4:44
136.4 C 224kPa 5:14
136.2 C 223kPa 5:44
136.4 C 225kPa 6:14
136.4 C 225kPa 6:44
136.2 C 221kPa 7:14
136.2 C 221kPa 7:14
VENTING 7:14
AIR DRYING 8:04
CYCLE COMPLETE 26:36

Digital Signature #
460694F6DFA3DC05
