

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3075

Name: RITTER 1  
Sterilizer ID: V2178198  
04/04/2023 02:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
2:00 199.0 F 4.0  
4:00 212.7 F 9.3  
6:00 226.0 F 7.9  
8:00 236.0 F 11.2  
10:00 235.6 F 8.9  
12:00 239.2 F 9.3  
14:00 249.5 F 14.9  
16:00 262.2 F 22.2  
18:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.2 F 28.6  
2:00 272.5 F 29.1  
3:00 271.8 F 28.3  
4:00 272.6 F 29.1  
  
Min 271.4 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 233.2 F 0.0  
10:00 232.9 F 0.0  
15:00 223.8 F 0.0  
20:00 222.9 F 0.0  
25:00 221.6 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:19  
CYCLE COMPLETE

04/04/2023 03:00 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_