
Midmark M11 - v1.0.5
Total Cycles: 464

Name: RITTER2
Sterilizer ID: V2313228
10/20/2021 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 152.4 F 2.3
4:00 157.3 F 5.5
6:00 186.5 F 10.4
8:00 208.3 F 5.2
10:00 224.8 F 10.3
12:00 228.1 F 7.7
14:00 241.4 F 13.1
16:00 241.5 F 11.7
18:00 254.1 F 18.1
20:00 265.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.8 F 30.0
2:00 271.8 F 29.2
3:00 272.0 F 29.9
4:00 272.5 F 29.8

Min 271.5 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 213.2 F 0.0
10:00 219.6 F 0.0
15:00 220.6 F 0.0
20:00 219.9 F 0.0
25:00 223.8 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

10/20/2021 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____