
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004875
12:58 20/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 1:59
STERILIZING 3:00
136.4 C 221kPa 3:00
136.1 C 217kPa 3:30
136.4 C 225kPa 4:00
136.4 C 223kPa 4:30
136.4 C 225kPa 5:00
136.4 C 222kPa 5:30
136.4 C 225kPa 6:00
136.4 C 222kPa 6:30
136.4 C 222kPa 6:30
VENTING 6:30
AIR DRYING 7:19
CYCLE COMPLETE 9:39

Digital Signature #
6A924837FBBF39C05
