

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3443

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2023 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
2:00 187.8 F 2.8  
4:00 200.7 F 7.1  
6:00 218.4 F 12.5  
8:00 228.5 F 8.0  
10:00 241.1 F 13.4  
12:00 240.1 F 11.2  
14:00 238.0 F 9.6  
16:00 252.1 F 16.5  
18:00 263.8 F 23.3  
20:00 271.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.3  
3:00 272.2 F 29.1  
4:00 272.4 F 28.9  
  
Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 233.7 F 0.0  
10:00 226.8 F 0.0  
15:00 228.7 F 0.0  
20:00 224.5 F 0.0  
25:00 224.1 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

06/26/2023 11:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_