
Midmark M11 - v1.0.5
Total Cycles: 1338

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 179.0 F 0.2
4:00 200.7 F 2.6
6:00 220.9 F 7.4
8:00 224.3 F 12.5
10:00 232.6 F 9.1
12:00 235.1 F 9.8
14:00 243.8 F 13.4
16:00 243.9 F 12.8
18:00 257.1 F 19.7
20:00 268.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.5 F 29.7
2:00 271.6 F 29.1
3:00 272.5 F 29.8
4:00 272.0 F 29.2

Min 271.5 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 234.7 F 0.0
10:00 226.4 F 0.0
15:00 224.2 F 0.0
20:00 221.7 F 0.0
25:00 221.2 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

07/25/2022 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____