

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4353

Name: RITTER 1  
Sterilizer ID: V2178198  
03/01/2024 02:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.7 F 0.0  
2:00 104.4 F 0.7  
4:00 122.4 F 2.5  
6:00 156.0 F 5.3  
8:00 181.2 F 8.8  
10:00 204.9 F 12.4  
12:00 220.4 F 4.3  
14:00 228.3 F 7.7  
16:00 237.4 F 11.5  
18:00 229.7 F 6.5  
20:00 241.5 F 11.4  
22:00 236.3 F 8.6  
24:00 248.6 F 14.3  
26:00 258.9 F 19.9  
28:00 268.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.4  
1:00 272.6 F 28.7  
2:00 272.5 F 28.9  
3:00 271.9 F 28.9  
4:00 271.3 F 28.4

Min 271.3 F 28.3  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
5:00 214.4 F 0.0  
10:00 211.1 F 0.0  
15:00 218.4 F 0.0  
20:00 215.6 F 0.0  
25:00 211.6 F 0.0  
30:00 207.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 29:51  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:08:36  
CYCLE COMPLETE

03/01/2024 03:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_