
Midmark M11 - v1.0.0
Total Cycles: 3546

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2023 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 73.1 F 0.0
4:00 80.2 F 0.0
6:00 111.7 F 0.1
8:00 143.1 F 1.3
10:00 167.6 F 3.1
12:00 186.1 F 5.7
14:00 206.6 F 8.9
16:00 224.4 F 12.4
18:00 224.9 F 5.7
20:00 234.5 F 9.7
22:00 240.1 F 11.0
24:00 239.7 F 10.8
26:00 235.1 F 8.2
28:00 249.0 F 14.5
30:00 259.8 F 20.5
32:00 269.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.6
2:00 272.5 F 29.0
3:00 271.5 F 28.4
4:00 272.4 F 28.8

Min 271.4 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 241.1 F 0.0
10:00 203.5 F 0.0
15:00 218.6 F 0.0
20:00 190.3 F 0.0
25:00 201.4 F 0.0
30:00 202.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:43
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:11:47
CYCLE COMPLETE

07/19/2023 09:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____