

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3662

Name: RITTER2  
Sterilizer ID: V2313228  
06/03/2024 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 92.8 F 0.0  
2:00 93.4 F 0.2  
4:00 103.4 F 1.3  
6:00 126.8 F 3.6  
8:00 162.7 F 6.5  
10:00 189.9 F 9.5  
12:00 196.1 F 13.4  
14:00 214.0 F 5.6  
16:00 221.1 F 9.1  
18:00 233.1 F 13.0  
20:00 232.8 F 9.1  
22:00 241.2 F 11.9  
24:00 242.3 F 12.0  
26:00 252.8 F 17.4  
28:00 262.7 F 23.1  
30:00 271.9 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 271.8 F 29.3  
2:00 272.2 F 29.7  
3:00 271.5 F 29.6  
4:00 271.6 F 29.2

Min 271.2 F 29.1  
Max 272.4 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
5:00 215.8 F 0.0  
10:00 193.1 F 0.0  
15:00 190.2 F 0.0  
20:00 186.2 F 0.0  
25:00 184.0 F 0.0  
30:00 184.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:30  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:08:56  
CYCLE COMPLETE

06/03/2024 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_