
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000809
14:26 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:42
PRESSURIZING 6:29
STERILIZING 7:59
136.0 C 220kPa 7:59
136.3 C 222kPa 8:29
136.0 C 220kPa 8:59
136.4 C 223kPa 9:29
136.0 C 219kPa 9:59
136.0 C 220kPa 10:29
136.5 C 223kPa 10:59
136.2 C 220kPa 11:29
136.2 C 220kPa 11:29
VENTING 11:29
AIR DRYING 12:33
CYCLE COMPLETE 17:20

Digital Signature #
361E84B19DCDA300
