
Midmark M11 - v1.0.5
Total Cycles: 1402

Name: RITTER2
Sterilizer ID: V2313228
08/23/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 82.5 F 0.0
4:00 83.6 F 0.0
6:00 109.2 F 0.0
8:00 159.2 F 2.2
10:00 186.6 F 6.1
12:00 203.8 F 10.5
14:00 228.1 F 5.9
16:00 228.3 F 8.7
18:00 237.9 F 13.1
20:00 237.7 F 10.3
22:00 236.8 F 8.6
24:00 249.3 F 15.1
26:00 260.9 F 21.8
28:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.8 F 29.0
2:00 271.9 F 29.2
3:00 272.2 F 29.3
4:00 272.1 F 29.2

Min 271.7 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 231.4 F 0.0
5:00 232.8 F 0.0
10:00 228.0 F 0.0
Dry Cycle Aborted

FILLING: 0:52
HEATING: 28:28
STERILIZING: 4:00
VENTING: 2:43
DRYING: 14:02
TOTAL CYCLE: 00:50:04

08/23/2022 10:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____