
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001234
15:07 29/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:57
PRESSURIZING 3:57
STERILIZING 5:06
136.0 C 220kPa 5:06
136.3 C 221kPa 5:37
136.4 C 225kPa 6:07
136.2 C 224kPa 6:37
136.4 C 225kPa 7:07
136.3 C 224kPa 7:37
136.3 C 224kPa 8:07
136.2 C 220kPa 8:37
136.2 C 220kPa 8:37
VENTING 8:37
AIR DRYING 9:32
CYCLE COMPLETE 19:41

Digital Signature #
176FA8DD55A37C29
