
Midmark M11 - v1.0.5
Total Cycles: 2706

Name: RITTER2
Sterilizer ID: V2313228
07/26/2023 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.7 F 0.0
2:00 121.3 F 0.0
4:00 151.1 F 0.2
6:00 177.6 F 3.3
8:00 194.4 F 7.1
10:00 209.8 F 11.1
12:00 222.5 F 6.1
14:00 237.9 F 11.0
16:00 230.9 F 7.7
18:00 245.8 F 13.2
20:00 243.9 F 12.4
22:00 255.6 F 18.5
24:00 266.1 F 24.9
26:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 272.6 F 29.8
2:00 271.6 F 29.3
3:00 272.6 F 29.5
4:00 272.1 F 29.6

Min 271.6 F 29.1
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 231.1 F 0.0
10:00 210.3 F 0.0
15:00 188.5 F 0.0
20:00 181.9 F 0.0
25:00 188.5 F 0.0
30:00 186.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:04:40
CYCLE COMPLETE

07/26/2023 03:15 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____