

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2520

Name: RITTER2  
Sterilizer ID: V2313228  
06/15/2023 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
2:00 176.6 F 3.1  
4:00 195.2 F 7.8  
6:00 216.9 F 13.4  
8:00 226.9 F 9.0  
10:00 236.2 F 10.1  
12:00 242.0 F 12.7  
14:00 241.7 F 11.8  
16:00 254.9 F 18.6  
18:00 266.5 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.6 F 29.8  
2:00 271.6 F 29.0  
3:00 272.5 F 30.0  
4:00 272.2 F 29.4  
  
Min 271.4 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 228.4 F 0.0  
10:00 223.3 F 0.0  
15:00 222.6 F 0.0  
20:00 220.5 F 0.0  
25:00 221.5 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 19:35  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

06/15/2023 10:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_