

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002669  
11:35 08/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:19  
STERILIZING 4:26  
136.4 C 223kPa 4:26  
136.4 C 226kPa 4:56  
136.4 C 223kPa 5:26  
136.2 C 223kPa 5:56  
136.4 C 225kPa 6:26  
136.4 C 225kPa 6:56  
136.1 C 224kPa 7:26  
136.2 C 224kPa 7:56  
136.2 C 224kPa 7:56  
VENTING 7:56  
AIR DRYING 8:46  
CYCLE COMPLETE 9:09

-----  
Digital Signature #  
CA1F2F49D324ADAF  
-----