
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000252
11:01 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:33
STERILIZING 4:32
136.0 C 220kPa 4:32
136.4 C 222kPa 5:02
136.3 C 224kPa 5:32
136.1 C 222kPa 6:02
136.4 C 224kPa 6:32
136.4 C 224kPa 7:02
136.4 C 222kPa 7:32
136.4 C 224kPa 8:02
136.4 C 224kPa 8:02
VENTING 8:02
AIR DRYING 8:58
CYCLE COMPLETE 9:07

Digital Signature #
F4BCFED42A408449
