

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001834  
10:35 12/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:59  
STERILIZING 3:02  
136.4 C 224kPa 3:02  
136.2 C 224kPa 3:32  
136.3 C 224kPa 4:02  
136.4 C 223kPa 4:32  
136.1 C 223kPa 5:02  
136.3 C 225kPa 5:32  
136.4 C 226kPa 6:02  
136.4 C 225kPa 6:32  
136.4 C 225kPa 6:32  
VENTING 6:32  
AIR DRYING 7:22  
CYCLE COMPLETE 16:06  
-----

Digital Signature #  
B8B4904C0E8CA09B  
-----