
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002348
14:41 02/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:51
136.5 C 222kPa 3:51
136.4 C 225kPa 4:21
136.4 C 224kPa 4:51
136.4 C 224kPa 5:21
136.1 C 222kPa 5:51
136.3 C 221kPa 6:21
135.9 C 220kPa 6:51
136.4 C 225kPa 7:21
136.4 C 225kPa 7:21
VENTING 7:21
AIR DRYING 8:12
CYCLE COMPLETE 8:25

Digital Signature #
34396F7C4342340C
