
Midmark M11 - v1.0.0
Total Cycles: 2982

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 167.7 F 3.3
4:00 185.8 F 7.6
6:00 210.4 F 12.7
8:00 219.7 F 7.1
10:00 234.8 F 12.3
12:00 235.2 F 9.2
14:00 239.9 F 9.4
16:00 247.9 F 14.1
18:00 260.3 F 21.0
20:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.7
2:00 271.8 F 28.7
3:00 272.1 F 28.5
4:00 271.9 F 28.7

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 231.9 F 0.0
10:00 235.8 F 0.0
15:00 228.9 F 0.0
20:00 223.9 F 0.0
25:00 226.9 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

03/08/2023 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____