
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000158
11:30 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:12
STERILIZING 5:37
136.4 C 220kPa 5:37
136.4 C 225kPa 6:07
136.2 C 223kPa 6:37
136.1 C 223kPa 7:07
136.2 C 223kPa 7:37
136.4 C 225kPa 8:07
136.4 C 226kPa 8:37
136.4 C 225kPa 9:07
136.4 C 225kPa 9:07
VENTING 9:07
AIR DRYING 10:09
CYCLE COMPLETE 14:32

Digital Signature #
DB3D04F514A1D865
