

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 000839  
15:18 16/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 2:59  
STERILIZING 4:06  
136.4 C 223kPa 4:06  
136.3 C 222kPa 4:36  
136.4 C 221kPa 5:06  
136.4 C 224kPa 5:36  
136.4 C 225kPa 6:06  
136.2 C 222kPa 6:36  
136.2 C 220kPa 7:06  
136.4 C 223kPa 7:36  
136.4 C 223kPa 7:36  
VENTING 7:36  
AIR DRYING 8:31  
CYCLE COMPLETE 8:35

-----  
Digital Signature #  
671EBE2A3EBFBFA0  
-----