
Midmark M11 - v1.0.0
Total Cycles: 5086

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 03:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.9 F 0.0
2:00 121.4 F 0.1
4:00 147.4 F 0.4
6:00 174.8 F 0.7
8:00 196.3 F 2.3
10:00 210.7 F 6.1
12:00 223.6 F 9.9
14:00 233.5 F 9.8
16:00 234.7 F 9.9
18:00 237.4 F 9.0
20:00 244.8 F 12.9
22:00 243.1 F 11.7
24:00 255.6 F 18.0
26:00 266.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.7
2:00 271.8 F 27.9
3:00 272.8 F 28.8
4:00 272.2 F 28.3

Min 271.6 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 224.6 F 0.0
10:00 222.4 F 0.0
15:00 218.5 F 0.0
20:00 217.4 F 0.0
25:00 233.8 F 0.0
30:00 235.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:36
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:15
CYCLE COMPLETE

11/13/2024 04:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____