
Midmark M11 - v1.0.0
Total Cycles: 1965

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 172.3 F 4.2
4:00 185.7 F 8.5
6:00 203.6 F 12.0
8:00 225.2 F 7.9
10:00 238.6 F 13.4
12:00 238.5 F 10.7
14:00 235.5 F 8.7
16:00 250.1 F 15.6
18:00 262.0 F 22.4
20:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.3
2:00 271.5 F 28.8
3:00 272.5 F 29.3
4:00 271.6 F 28.6

Min 271.4 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 214.0 F 0.0
10:00 244.3 F 0.0
15:00 243.4 F 0.0
20:00 230.2 F 0.0
25:00 221.5 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/27/2022 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____