
Midmark M11 - v1.0.5
Total Cycles: 103

Name: RITTER2
Sterilizer ID: V2313228
06/15/2021 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 187.7 F 3.8
4:00 203.4 F 9.0
6:00 218.7 F 8.2
8:00 229.4 F 10.0
10:00 233.8 F 8.6
12:00 241.7 F 11.5
14:00 247.0 F 14.1
16:00 259.2 F 21.0
18:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.4 F 29.5
2:00 271.7 F 29.4
3:00 272.2 F 29.2
4:00 272.4 F 29.7

Min 271.4 F 28.9
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 239.6 F 0.0
10:00 211.3 F 0.0
15:00 200.0 F 0.0
20:00 188.1 F 0.0
25:00 184.1 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

06/15/2021 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____