

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003229  
11:03 28/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:55  
STERILIZING 2:44  
136.0 C 220kPa 2:44  
136.4 C 221kPa 3:14  
136.3 C 220kPa 3:44  
136.4 C 224kPa 4:14  
136.2 C 220kPa 4:44  
136.4 C 223kPa 5:14  
136.4 C 224kPa 5:44  
136.4 C 221kPa 6:14  
136.4 C 221kPa 6:14  
VENTING 6:14  
AIR DRYING 7:08  
CYCLE COMPLETE 7:20

-----  
Digital Signature #  
B6A1E6EF2C60DDB6  
-----