
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001239
16:46 31/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:33
STERILIZING 4:37
136.0 C 220kPa 4:37
136.2 C 223kPa 5:07
136.2 C 222kPa 5:37
136.2 C 223kPa 6:07
136.2 C 224kPa 6:37
136.1 C 223kPa 7:07
136.4 C 224kPa 7:37
136.4 C 225kPa 8:07
136.4 C 225kPa 8:07
VENTING 8:08
AIR DRYING 9:00
CYCLE COMPLETE 11:24

Digital Signature #
4D19F96869C6F7C4
