

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1509

Name: RITTER 1  
Sterilizer ID: V2178198  
01/11/2022 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.0 F 0.0  
2:00 174.0 F 3.2  
4:00 175.8 F 7.1  
6:00 197.3 F 11.7  
8:00 216.9 F 6.1  
10:00 228.1 F 11.2  
12:00 231.0 F 7.4  
14:00 243.1 F 13.1  
16:00 241.9 F 11.2  
18:00 254.1 F 17.6  
20:00 265.3 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.0  
2:00 271.7 F 28.6  
3:00 272.3 F 28.9  
4:00 271.6 F 28.6  
  
Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 233.1 F 0.0  
10:00 235.3 F 0.0  
15:00 233.3 F 0.0  
20:00 229.2 F 0.0  
25:00 230.5 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

01/11/2022 12:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_