
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001572
16:47 11/10/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:54
PRESSURIZING 6:35
STERILIZING 8:06
136.0 C 220kPa 8:06
136.1 C 219kPa 8:36
136.2 C 221kPa 9:06
136.4 C 223kPa 9:37
136.4 C 218kPa 10:07
136.3 C 222kPa 10:37
136.2 C 220kPa 11:07
135.8 C 219kPa 11:36
VENTING 11:37
AIR DRYING 12:40
CYCLE COMPLETE 22:50

Digital Signature #
3AB095D9F691A715
