
Midmark M11 - v1.0.0
Total Cycles: 4441

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2024 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 72.8 F 0.2
4:00 75.6 F 1.1
6:00 100.4 F 3.1
8:00 134.2 F 5.7
10:00 160.2 F 8.8
12:00 181.1 F 12.1
14:00 203.9 F 2.6
16:00 209.7 F 5.7
18:00 220.1 F 8.9
20:00 230.5 F 12.7
22:00 232.0 F 8.3
24:00 243.1 F 13.2
26:00 241.2 F 11.1
28:00 252.7 F 16.9
30:00 263.0 F 22.8
32:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.4
3:00 272.5 F 29.1
4:00 271.6 F 28.4

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 220.0 F 0.0
10:00 232.0 F 0.0
15:00 228.5 F 0.0
20:00 228.4 F 0.0
25:00 227.5 F 0.0
30:00 227.2 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:23
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:10:52
CYCLE COMPLETE

04/10/2024 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____