
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005998
9:36 18/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:31
136.4 C 221kPa 3:31
136.3 C 224kPa 4:01
136.4 C 225kPa 4:31
136.4 C 222kPa 5:01
136.4 C 220kPa 5:31
136.4 C 221kPa 6:01
136.4 C 222kPa 6:31
136.4 C 221kPa 7:01
136.4 C 221kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 8:00

Digital Signature #
36F8FCA0FC7F31A4
