

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3426

Name: RITTER 1  
Sterilizer ID: V2178198  
06/22/2023 07:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.5 F 0.0  
2:00 76.9 F 0.0  
4:00 82.8 F 0.0  
6:00 109.2 F 0.0  
8:00 142.5 F 1.4  
10:00 173.1 F 4.0  
12:00 189.7 F 7.1  
14:00 203.2 F 10.6  
16:00 220.3 F 6.8  
18:00 224.6 F 7.3  
20:00 234.0 F 11.3  
22:00 231.7 F 7.3  
24:00 244.7 F 13.0  
26:00 241.9 F 11.2  
28:00 253.9 F 17.3  
30:00 264.5 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.0 F 28.8  
2:00 272.3 F 28.6  
3:00 271.6 F 28.5  
4:00 273.0 F 29.0  
  
Min 271.3 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
5:00 209.1 F 0.0  
10:00 230.9 F 0.0  
15:00 225.9 F 0.0  
20:00 220.1 F 0.0  
25:00 218.3 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:56  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

06/22/2023 08:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_