
Midmark M11 - v1.0.0
Total Cycles: 3167

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2023 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.6 F 0.0
2:00 197.6 F 2.6
4:00 194.0 F 6.6
6:00 210.1 F 11.8
8:00 221.0 F 6.6
10:00 234.6 F 11.8
12:00 235.3 F 8.9
14:00 239.2 F 9.8
16:00 247.3 F 13.9
18:00 259.7 F 20.6
20:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.3
3:00 272.3 F 29.0
4:00 272.4 F 28.8

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 215.4 F 0.0
10:00 232.8 F 0.0
15:00 239.1 F 0.0
20:00 231.2 F 0.0
25:00 219.6 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

04/27/2023 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____