
Midmark M11 - v1.0.5
Total Cycles: 3055

Name: RITTER2
Sterilizer ID: V2313228
10/27/2023 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	68.2 F	0.0
2:00	80.0 F	0.0
4:00	125.6 F	0.0
6:00	104.6 F	0.1
8:00	142.3 F	0.3
10:00	170.8 F	0.3
12:00	195.2 F	0.7
14:00	212.2 F	1.3
16:00	218.0 F	2.1
18:00	220.6 F	2.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 221F
CHAMBER PRESS 3.0PSI
** ITEMS NOT STERILE **

10/27/2023 07:53 AM