
Midmark M11 - v1.0.0
Total Cycles: 4270

Name: RITTER 1
Sterilizer ID: V2178198
02/05/2024 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.3 F 0.0
2:00 122.1 F 1.2
4:00 130.7 F 3.5
6:00 155.4 F 6.4
8:00 178.0 F 9.8
10:00 198.5 F 11.8
12:00 208.6 F 5.6
14:00 220.1 F 9.6
16:00 233.7 F 10.9
18:00 236.2 F 9.9
20:00 237.5 F 9.4
22:00 245.5 F 13.1
24:00 256.5 F 18.9
26:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.1 F 29.2
2:00 272.5 F 29.1
3:00 272.1 F 28.9
4:00 272.6 F 29.5

Min 270.9 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 213.7 F 0.0
10:00 225.7 F 0.0
15:00 224.0 F 0.0
20:00 231.2 F 0.0
25:00 227.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:47
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:30
CYCLE COMPLETE

02/05/2024 03:01 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____