
Midmark M11 - v1.0.5
Total Cycles: 3179

Name: RITTER2
Sterilizer ID: V2313228
12/04/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 162.4 F 1.3
4:00 167.7 F 4.4
6:00 196.5 F 8.9
8:00 211.6 F 11.8
10:00 224.1 F 7.7
12:00 235.2 F 12.6
14:00 235.7 F 10.0
16:00 238.5 F 9.0
18:00 248.3 F 15.1
20:00 259.9 F 21.5
22:00 270.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.9 F 30.0
2:00 271.8 F 29.3
3:00 271.5 F 29.7
4:00 272.9 F 29.9

Min 271.1 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 204.2 F 0.0
10:00 201.2 F 0.0
15:00 197.2 F 0.0
20:00 195.9 F 0.0
25:00 195.2 F 0.0
30:00 195.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

12/04/2023 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____