
Midmark M11 - v1.0.5
Total Cycles: 3689

Name: RITTER2
Sterilizer ID: V2313228
06/12/2024 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 148.5 F 1.9
4:00 160.3 F 4.9
6:00 183.7 F 8.6
8:00 204.5 F 12.8
10:00 221.4 F 6.1
12:00 228.3 F 10.4
14:00 234.5 F 8.8
16:00 239.2 F 12.2
18:00 239.6 F 10.6
20:00 251.3 F 16.5
22:00 261.7 F 22.4
24:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 271.4 F 29.4
2:00 271.4 F 29.4
3:00 271.8 F 29.4
4:00 272.2 F 29.6

Min 271.2 F 29.4
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 190.0 F 0.0
10:00 191.1 F 0.0
15:00 188.3 F 0.0
20:00 184.4 F 0.0
25:00 181.4 F 0.0
30:00 179.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:16
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:04
CYCLE COMPLETE

06/12/2024 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____