

-----  
Midmark M11 - v1.0.5  
Total Cycles: 70

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2021 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
2:00 189.9 F 4.1  
4:00 196.9 F 9.7  
6:00 217.6 F 6.3  
8:00 230.4 F 11.2  
10:00 236.5 F 9.5  
12:00 237.9 F 9.3  
14:00 251.0 F 16.1  
16:00 263.5 F 23.5  
18:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.3 F 29.5  
2:00 271.4 F 28.9  
3:00 272.6 F 29.6  
4:00 271.9 F 29.0  
  
Min 271.3 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 225.9 F 0.0  
10:00 236.3 F 0.0  
15:00 202.4 F 0.0  
20:00 170.4 F 0.0  
25:00 157.5 F 0.0  
Dry Cycle Aborted

FILLING: 2:02  
HEATING: 18:01  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 25:37  
TOTAL CYCLE: 00:52:45

06/07/2021 12:01 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01