
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002836
10:57 14/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:36
136.4 C 223kPa 4:36
136.0 C 222kPa 5:06
136.4 C 221kPa 5:36
136.4 C 223kPa 6:06
136.4 C 223kPa 6:36
136.3 C 222kPa 7:06
136.2 C 222kPa 7:36
136.3 C 221kPa 8:06
136.3 C 221kPa 8:06
VENTING 8:06
AIR DRYING 9:02
CYCLE COMPLETE 9:03

Digital Signature #
C2CDAE4D1D1D0AC7
