
Midmark M11 - v1.0.0
Total Cycles: 739

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2021 06:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 75.3 F 0.0
4:00 122.2 F 0.1
6:00 123.9 F 0.1
8:00 151.9 F 0.2
10:00 187.0 F 0.4
12:00 205.3 F 4.0
14:00 218.1 F 7.1
16:00 227.0 F 10.6
18:00 231.6 F 8.0
20:00 238.7 F 11.2
22:00 235.1 F 8.6
24:00 239.8 F 9.9
26:00 247.8 F 14.0
28:00 260.6 F 21.1
30:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.1 F 28.4
2:00 272.6 F 29.0
3:00 271.5 F 28.3
4:00 272.2 F 28.5

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 221.2 F 0.0
10:00 224.7 F 0.0
15:00 214.7 F 0.0
20:00 243.8 F 0.0
25:00 202.5 F 0.0
30:00 181.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:29
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:08:12
CYCLE COMPLETE

03/24/2021 07:34 AM
ACCEPTED: X
REJECTED:
Operator: U01