
Midmark M11 - v1.0.0
Total Cycles: 1213

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2021 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.3 F 0.0
2:00 104.2 F 0.0
4:00 124.5 F 0.1
6:00 171.7 F 0.2
8:00 184.0 F 2.0
10:00 204.0 F 5.1
12:00 215.9 F 8.4
14:00 227.6 F 12.2
16:00 231.8 F 7.6
18:00 242.3 F 12.4
20:00 239.0 F 9.7
22:00 238.1 F 9.2
24:00 248.6 F 14.2
26:00 259.9 F 20.4
28:00 270.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.7 F 28.7
2:00 271.7 F 27.9
3:00 272.1 F 28.6
4:00 272.4 F 28.3

Min 271.1 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 235.3 F 0.0
10:00 223.1 F 0.0
15:00 213.2 F 0.0
20:00 213.7 F 0.0
25:00 224.3 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:57
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:07:15
CYCLE COMPLETE

09/28/2021 02:37 PM
ACCEPTED: X
REJECTED:
Operator: U01