
Midmark M11 - v1.0.5
Total Cycles: 3501

Name: RITTER2
Sterilizer ID: V2313228
04/05/2024 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.0 F 0.0
2:00 111.8 F 1.1
4:00 135.6 F 3.5
6:00 161.9 F 6.8
8:00 184.8 F 10.8
10:00 213.0 F 4.7
12:00 228.3 F 8.6
14:00 238.8 F 13.4
16:00 233.8 F 10.2
18:00 235.1 F 8.6
20:00 247.3 F 14.8
22:00 259.0 F 21.3
24:00 269.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 30.4
2:00 271.4 F 29.7
3:00 272.6 F 30.0
4:00 272.2 F 30.3

Min 270.9 F 29.2
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 232.2 F 0.0
10:00 218.3 F 0.0
15:00 185.2 F 0.0
20:00 179.5 F 0.0
25:00 175.7 F 0.0
30:00 173.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

04/05/2024 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____