
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002078
9:17 23/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:05
136.0 C 221kPa 4:05
136.4 C 226kPa 4:35
136.4 C 226kPa 5:05
136.3 C 224kPa 5:35
136.3 C 225kPa 6:05
136.4 C 225kPa 6:35
136.4 C 220kPa 7:05
136.1 C 222kPa 7:35
136.1 C 222kPa 7:35
VENTING 7:35
AIR DRYING 8:34
CYCLE COMPLETE 8:40

Digital Signature #
E963F8CDB8327B01
