
Midmark M11 - v1.0.0
Total Cycles: 1902

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2022 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.7 F 0.0
2:00 129.1 F 0.0
4:00 153.5 F 1.2
6:00 184.4 F 4.5
8:00 202.4 F 8.5
10:00 221.7 F 12.9
12:00 225.7 F 7.0
14:00 237.5 F 11.7
16:00 235.3 F 8.5
18:00 241.5 F 10.6
20:00 246.2 F 13.1
22:00 258.6 F 19.8
24:00 269.6 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.6 F 28.7
2:00 271.8 F 28.3
3:00 272.5 F 28.7
4:00 271.6 F 28.1

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 189.9 F 0.0
10:00 196.5 F 0.0
15:00 201.0 F 0.0
20:00 205.1 F 0.0
25:00 214.1 F 0.0
30:00 206.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

05/03/2022 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____