
Midmark M11 - v1.0.0
Total Cycles: 3824

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 175.2 F 2.9
4:00 191.6 F 7.1
6:00 209.0 F 12.0
8:00 224.6 F 6.7
10:00 237.3 F 11.6
12:00 235.6 F 8.6
14:00 241.1 F 10.7
16:00 246.6 F 13.2
18:00 258.5 F 19.5
20:00 269.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.6 F 28.4
2:00 271.5 F 28.3
3:00 271.7 F 28.3
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 225.2 F 0.0
10:00 241.4 F 0.0
15:00 231.6 F 0.0
20:00 225.0 F 0.0
25:00 228.9 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

09/26/2023 05:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____